

Sl#	Material Description	Quantity
1	ANALOG ROCKWELL HARDNESS TESTING MACHINE.	1 No.

SPECIFICATION:

1. HARDNESS DISPLAY: ANALOG
2. RESOLUTION: 0.5 HR GRADUATION.
3. PRELIMINARY TEST FORCE: AUTOMATIC PRESENTING DIAL GUAGE.
4. TOTAL TEST FORCE SETTING: WEIGHT CHANGE.
5. TOTAL TEST FORCE CONTROL: MANUAL.
6. TEST FORCE DURATION: MANUAL.
7. POWER SUPPLY: NO POWER REQUIRED.
8. STANDARE: AS PER STANDARD ASTM E18.
9. MODEL : HR-110R ORDER NO. : 963-210 OF MAKE MUTUTOYO OR MODEL: RAS OF MAKE: SAROJ OR EQUIVALENT.
10. THE BASIC OPERATION IS ALL MANUFAL INCLUDING WEIGHT CHANGING (TOTAL TEST SELECTION).
11. THE MACHINE IS CAPABLE OF MEASURING THE HARDNESS IN ALL SCALE OF ROCKWELL HARDNESS.

NOTE:

1. Bidder Should submit their offer as per terms & conditions given under Annexure 1 2 & 3[ANNEXURE -1(Pre -Qualification Part) ANNEXURE-2(Techno-commercial Cum Price Part) & Annexure -3 (Pre-Qualification Criteria)]
2. Offer submitted in single part will be rejected.
3. INSTALLATION OF ANALOG ROCKWELL HARDNESS TESTING MACHINE IS UNDER THE SCOPE OF SUPPLIER. SUCESSFUL BIDDER OR SUPPLIER HAS TO INSTALL THE MACHINE AT UCIL-TURAMDIH
4. BIDDER MUST SUBMIT DETAILED BROCHURE WITH SPECIFICATION OF OFFERED PRODUCT ALONG WITH OFFER.
5. BIDDER MUST SPECIFY MAKE OF OFFERED PRODUCT (ANALOG ROCKWELL HARDNESS TESTING MACHINE) ALONG WITH OFFER.

(AKHIL RAJ)
DY. SUPDT. (MECH. / PURCHASE)